

PLEASE NOTE: Due to regular software updates, please visit the download page on our website etiltd.com/downloads to ensure you have the latest version



THERMADATA® LITE

Temperature recording thermometers



Manufactured by
Electronic Temperature Instruments Ltd
Worthing · West Sussex · BN14 8HQ
01903 202151 · sales@etiltd.com · etiltd.com



545-310/03.01.19

Quick start guide

PC SOFTWARE FOR THIS INSTRUMENT

Install ThermaData Studio which can be downloaded free from etilt.com/downloads. For further queries, please contact our technical sales on 01903 202151 or technical@etilt.com

SETTING UP THE LOGGER

- Connect the Instrument to the PC using a USB C lead.
- Select the 'ThermaData@' tab from the top menu bar.
- Select 'Instrument settings' option.
- Select the correct serial number from the drop down list.
- Fill in an optional instrument name and select the Over/Under alarm levels.
- Select a reading interval rate.
- Select a Start method and fill in the options associated with it.
- Select a Stop method and fill in the options associated with it (Stop at time can only be used if 'Start At Time' is used).
- Tick or untick the square box for 'Stop Logging on Button Hold'.
- Fill in the optional 'Instructions' that will appear on the PDF report.
- Click the 'Send Settings' button and a confirmation window will appear.
- Start the logger by disconnecting the instrument from the PC and pushing the button ('Start at Time' requires no button push). 4 red and green LED flashes will indicate the start method has been activated.
- Pressing the button now will give you 3 red or 3 green LED flashes depending if an alarm has been activated. If the Instrument has not yet met the start criteria there will be 2 alternating red and green LED flashes.

READING THE LOGGER

- Connect the Instrument to the PC using the USB C lead.
- A PDF log summary will automatically pop up as a Mass Storage Device.
- Alternatively, using ThermaData Studio select the 'Download' option found in the 'Home' tab and select the correct serial number from the drop down list.
- To view individual readings select the 'Data' option found in the 'Home' tab.

BATTERY CHANGE

This instrument is fitted with a Lithium ½ AA 3.6V non-rechargeable Green Energy battery. Failure to use the same type of battery could result in errors with the instrument.

Once a new battery is fitted you must go to ThermaData Studio, select the 'ThermaData@' tab and select the 'Factory Reset' option, selecting the correct serial number from the drop down list.

This product uses a non-recharge battery.

For a detailed explanation on functionality and LED configurations please go to ThermaData Studio and select the 'Help' tab in the top right corner.